


### B1651-E5 Summary

The test data was installed into a spreadsheet calculation curve with a recommended maximum operating temperature of 150°C. At a 2-amp continuous input, the test probes temperature rise was 53°C, ambient temperature cannot exceed 97°C in order to maintain below the recommended maximum operating temperature of 150°C.



Current carrying capacity is not an absolute, but an application-dependent condition. Factors are: quantity/mounting/spacing, airflow, degradation factors, all affect the current rating. This is a representation of 1 contact. The designer must select appropriate current levels for these factors and its particular use, which varies, dependent on application. Therefore, testing is recommended for each particular application.



**RIKA DENSHI AMERICA, INC.**

112 Frank Mossberg Drive, Attleboro, MA 02703 USA

Tel: (508) 226-2080 / Toll Free: (877) 902-TEST / Fax: (508) 226-2980

Website: [www.testprobe.com](http://www.testprobe.com)